

Display Report

Analysis Info

Analysis Name D:\Data\test\0829\3,6-erjiezixianjizhetang.d
Method esi-lcms-zhengfu.m
Sample Name 3,6-erjiezixianjizhetang
Comment

Operator
Instrument bruger amaZon SL

Acquisition Parameter

| | | | | | |
|-------------------|--------------|--------------|-----------|--------------------------|----------|
| Ion Source Type | ESI | Ion Polarity | Positive | Alternating Ion Polarity | on |
| Mass Range Mode | UltraScan | Scan Begin | 70 m/z | Scan End | 1000 m/z |
| Capillary Exit | 140.0 V | n/a | n/a | Trap Drive | 50.6 |
| Accumulation Time | 3508 μ s | Averages | 5 Spectra | Auto MS/MS | off |

