

# Display Report

## Analysis Info

Analysis Name D:\Data\TEST\1107\yixianzicaosu\_84\_01\_324.d  
Method 324.m  
Sample Name yixianzicaosu  
Comment

Instrument amaZon SL

## Acquisition Parameter

Ion Source Type	ESI	Ion Polarity	Negative	Alternating Ion Polarity	on
Mass Range Mode	UltraScan	Scan Begin	70 m/z	Scan End	1000 m/z
Capillary Exit	-140.0 V	n/a	n/a	Trap Drive	79.4
Accumulation Time	122560 $\mu$ s	Averages	5 Spectra	Auto MS/MS	off

